

## CERTIFICATE OF CALIBRATION

Company: TEXAS A&M UNIVERSITY

This document certifies that the below listed product has been calibrated to manufacturer's standards using Certified Reference Materials and Reference Materials.

Model: TRACER 5g  
Serial No: 900G7838  
Date Calibrated: AUGUST 21, 2020

Certified by:



# Tracer 5 Precious Metals Check Sample Report



Date of Report: August 19, 2020

Instrument Serial Number: 900G7838

Sample: Silver Round

Measurement	Cu	Ag
1	6.89	93.11
2	6.85	93.15
3	6.87	93.13
4	6.86	93.14
5	6.88	93.12
6	6.84	93.16
7	6.85	93.15
8	6.87	93.13
9	6.86	93.14
10	6.86	93.15



Average Measured Value	6.862	93.138
Standard Deviation	0.017	0.017
Acceptance Limit MIN	6.777	93.052
Acceptance Limit MAX	6.948	93.223

General instrument performance may be checked by taking a measurement of the supplied silver sample. The measured value for each element should be within the acceptance limits shown above. If not, contact Bruker for support. Email: [service.hmp@bruker.com](mailto:service.hmp@bruker.com).

Measurement values are in Weight Percent (%)

\*Use the following configuration to check instrument performance:

APPLICATION:	Precious Metals 2
METHOD:	PrecMetals
SETTINGS:	15 seconds

# Tracer5 GeoExploration Check Sample Report



**Date of Report:** August 19, 2020  
**Instrument Serial Number:** 900G7226  
**Sample:** CS-M2

Measurement	Al2O3	K2O	Ca	Mn	Fe	Cu	Pb
1	10.27	4.63	0.60	0.105	1.61	0.0202	0.0816
2	10.11	4.62	0.59	0.107	1.62	0.0191	0.0806
3	10.50	4.63	0.60	0.108	1.60	0.0205	0.0813
4	10.07	4.63	0.60	0.107	1.62	0.0206	0.0786
5	10.29	4.65	0.59	0.108	1.60	0.0201	0.0818
6	10.40	4.64	0.59	0.105	1.62	0.0198	0.0825
7	10.37	4.64	0.60	0.102	1.62	0.0214	0.0788
8	10.24	4.62	0.60	0.107	1.61	0.0207	0.0791
9	10.52	4.61	0.60	0.104	1.61	0.0215	0.0818
10	10.35	4.61	0.61	0.106	1.61	0.0210	0.0809

<b>Average Measured Value</b>	<b>10.31</b>	<b>4.63</b>	<b>0.60</b>	<b>0.106</b>	<b>1.61</b>	<b>0.0205</b>	<b>0.0807</b>
<b>Standard Deviation</b>	0.15	0.01	0.01	0.002	0.01	0.0007	0.0014
<b>Acceptance Limit MIN</b>	<b>9.28</b>	<b>4.40</b>	<b>0.54</b>	<b>0.092</b>	<b>1.50</b>	<b>0.0154</b>	<b>0.0710</b>
<b>Acceptance Limit MAX</b>	<b>11.34</b>	<b>4.86</b>	<b>0.65</b>	<b>0.120</b>	<b>1.72</b>	<b>0.0256</b>	<b>0.0904</b>



*General instrument performance may be checked by taking a measurement of the supplied CS-M2 sample\*. The measured value for each element should be within the acceptance limits shown above. If not, contact Bruker for support. Email: [service.hmp@bruker.com](mailto:service.hmp@bruker.com).*

*Measurement values are in Weight Percent (%)*

*\*Use the following configuration to check instrument performance:*

APPLICATION:	GeoExploration
METHOD:	Oxide3phase
SETTINGS:	3 Phase : 20 seconds each



# Tracer 5 MudRock He Check Sample Report



Date of Report: August 19, 2020

Instrument Serial Number: 900G7838

Sample: MudRock

Measurement	Al	P	Ca	Fe	Cu	Zn	Rb	Sr
1	4.57	5.90	0.70	0.235	0.01	0.045	0.036	0.011
2	4.60	5.92	0.70	0.230	0.01	0.045	0.036	0.011
3	4.58	5.91	0.71	0.232	0.01	0.045	0.036	0.011
4	4.58	5.94	0.71	0.232	0.01	0.046	0.035	0.011
5	4.56	5.92	0.70	0.231	0.01	0.044	0.036	0.011
6	4.56	5.87	0.70	0.232	0.01	0.046	0.036	0.011
7	4.58	5.89	0.70	0.230	0.01	0.046	0.036	0.011
8	4.60	5.91	0.70	0.232	0.01	0.045	0.036	0.011
9	4.56	5.89	0.70	0.233	0.01	0.046	0.036	0.011
10	4.56	5.89	0.70	0.235	0.01	0.045	0.034	0.011



<b>Average Measured Value</b>	<b>4.576</b>	<b>5.90</b>	<b>0.70</b>	<b>0.232</b>	<b>0.010</b>	<b>0.045</b>	<b>0.036</b>	<b>0.011</b>
<b>Standard Deviation</b>	0.016	0.019	0.003	0.002	0.000	0.001	0.001	0.000
<b>Acceptance Limit MIN</b>	<b>4.039</b>	<b>5.22</b>	<b>0.61</b>	<b>0.192</b>	<b>0.005</b>	<b>0.036</b>	<b>0.026</b>	<b>0.008</b>
<b>Acceptance Limit MAX</b>	<b>5.113</b>	<b>6.59</b>	<b>0.80</b>	<b>0.273</b>	<b>0.015</b>	<b>0.055</b>	<b>0.045</b>	<b>0.014</b>

General instrument performance may be checked by taking a measurement of the supplied check sample\*. The measured value for each element should be within the acceptance limits shown above. If not, contact Bruker for support. Email: support.hmp@bruker.com.

Measurement values are in Weight Percent (%)

\*Use the following configuration to check instrument performance:

APPLICATION:	MudRock He**
METHOD:	MudRock He**
SETTINGS:	30 seconds Phase 1, 60 seconds Phase 2

\*\*Make sure the He flow set up is used for these measurements

# Tracer 5 Limestone Check Sample Report

**Date of Report:** December 6, 2022

**Instrument Serial Number:** 900G7838

**Sample:** Limestone

Measurement	Al2O3	SiO2	CaCO3	Fe2O3
1	0.167	0.282	99.461	0.063
2	0.156	0.318	99.442	0.059
3	0.159	0.306	99.441	0.065
4	0.160	0.281	99.463	0.064
5	0.152	0.302	99.459	0.060
6	0.167	0.288	99.450	0.063
7	0.167	0.297	99.441	0.063
8	0.165	0.296	99.455	0.058
9	0.148	0.278	99.476	0.065
10	0.162	0.312	99.444	0.057




<b>Average Measured Value</b>	<b>0.160</b>	<b>0.000</b>	<b>99.453</b>	<b>0.062</b>
<b>Standard Deviation</b>	0.007	0.000	0.012	0.003
<b>Acceptance Limit MIN</b>	<b>0.127</b>	<b>0.000</b>	<b>97.000</b>	<b>0.048</b>
<b>Acceptance Limit MAX</b>	<b>0.194</b>	<b>0.000</b>	<b>99.512</b>	<b>0.075</b>

General instrument performance may be checked by measuring the check sample as described below (Note: measurements must be made on the smooth surface). The sample should be centered on the window to avoid variation from heterogeneity. The measured value for each element should be within the acceptance limits shown above. If not, contact Bruker for support. Email: [service.hmp@bruker.com](mailto:service.hmp@bruker.com).

Measurement values are in Weight Percent (%)

\*Use the following configuration to check instrument performance:

APPLICATION:	Limestone
METHOD:	Limestone
SETTINGS:	Phase 1: 120 seconds



## CERTIFICATE OF CALIBRATION

Company: TEXAS A&M UNIVERSITY

This document certifies that the below listed product has been calibrated to manufacturer's standards using Certified Reference Materials and Reference Materials.

Model: TRACER 5G

Serial No: 900G7838

Date Calibrated: JUANUARY 9, 2022

Certified by:





# Tracer 5 MudRock Air Check Sample Report



Date of Report: January 9, 2022

Instrument Serial Number: 900G7838

Sample: MudRock

Measurement	Al	P	Ca	Fe	Cu	Zn	Rb	Sr
1	4.12	6.79	0.71	0.263	0.01	0.049	0.038	0.011
2	4.13	6.81	0.71	0.252	0.01	0.049	0.038	0.011
3	4.12	6.83	0.71	0.249	0.01	0.049	0.037	0.011
4	4.11	6.83	0.71	0.260	0.01	0.049	0.037	0.011
5	4.11	6.81	0.72	0.255	0.01	0.048	0.037	0.011
6	4.12	6.79	0.71	0.255	0.01	0.048	0.037	0.011
7	4.06	6.79	0.72	0.254	0.01	0.048	0.038	0.011
8	4.13	6.77	0.71	0.255	0.01	0.049	0.037	0.011
9	4.13	6.79	0.72	0.254	0.01	0.049	0.036	0.011
10	4.04	6.78	0.71	0.259	0.01	0.049	0.038	0.011



<b>Average Measured Value</b>	<b>4.107</b>	<b>6.80</b>	<b>0.71</b>	<b>0.255</b>	<b>0.010</b>	<b>0.049</b>	<b>0.037</b>	<b>0.011</b>
<b>Standard Deviation</b>	<b>0.032</b>	<b>0.020</b>	<b>0.002</b>	<b>0.004</b>	<b>0.000</b>	<b>0.000</b>	<b>0.000</b>	<b>0.000</b>
<b>Acceptance Limit MIN</b>	<b>3.539</b>	<b>6.02</b>	<b>0.62</b>	<b>0.189</b>	<b>0.004</b>	<b>0.040</b>	<b>0.029</b>	<b>0.008</b>
<b>Acceptance Limit MAX</b>	<b>4.676</b>	<b>7.58</b>	<b>0.81</b>	<b>0.322</b>	<b>0.016</b>	<b>0.057</b>	<b>0.046</b>	<b>0.014</b>

General instrument performance may be checked by taking ten measurements of the supplied check sample\* and then calculating the average measured values. The average measured value for each element should be within the acceptance limits shown above. If not, contact Bruker for support. Email: [support.hmp@bruker.com](mailto:support.hmp@bruker.com). The side with the "O" should face instrument.

Measurement values are in Weight Percent (%)

\*Use the following configuration to check instrument performance:

APPLICATION:	MudRock Air
METHOD:	MudRock Air
SETTINGS:	30 secons Phase 1, 60 seconds Phase 2